



The European Association of
Medical devices Notified Bodies

Team-NB Position Paper

Editor: **Team-NB** Adoption date 09/04/2025 Version 2

European Artificial Intelligence Act

1 **About Team-NB**

2 TEAM-NB (The European Association for Medical Devices of Notified Bodies) is an association
3 representing notified bodies (NBs) active in the certification of medical devices¹ under EU regulations.
4 Our main objectives are promoting high standards of conformity assessment, ensuring harmonised
5 implementation of the MDR/IVDR, and facilitating dialogue between NBs, regulators, industry, and
6 other stakeholders.

7 **Background**

8 The European Commission Artificial Intelligence (AI) Regulation (Regulation (EU) 2024/1689), in the
9 following “AI Act” or “AIA” has entered into force on 01 August 2024. The AI Act is designed horizontally
10 to cover most business sectors and introduces a risk-based approach to determine the level of scrutiny
11 on AI systems being placed on the Union market or being used in the Union. Use of AI systems as a
12 safety component of or as a medical device or *in vitro* diagnostic² (MDAI) is considered high-risk
13 according to the AI Regulation, if a conformity assessment would be required according to risk
14 classification under Medical Device Regulation (EU) 2017/745 (MDR) or In Vitro Diagnostics Regulation
15 (EU) 2017/746 (IVDR) (AIA Art. 6(1)).

16 According to Article 113 (c) (AIA), the high-risk scope under Article 6(1) will become applicable on 02
17 August 2027. From that date onwards, newly introduced or legacy medical devices undergoing a
18 significant change (Art. 111 (2)) and fulfilling the aforementioned conditions (i.e. requirement for third
19 party conformity assessment and use as safety component or as a device itself) will have to comply
20 with the provisions of the AI Act.

21 The members of Team-NB publish this position to express their opinion regarding the implementation
22 and application of the EU AI Act, with a focus on the interface and overlap with MDR and IVDR.

¹ If not indicated otherwise, the term ‘medical device’ includes ‘*in vitro* diagnostic medical device’.

² For the purpose of this document, an ‘AI system in or as medical devices or *in vitro* diagnostics’ is ‘medical device AI’ or ‘MDAI’.



The European Association of
Medical devices Notified Bodies

Team-NB Position Paper

24 **Summary**

25 This position paper provides an overview of the Team-NB perspective on the challenges of the AI Act
26 with particular attention to its implementation.

27 Because of the tight timelines for implementation, we welcome a coordinated approach between
28 member states for the designation and oversight of NBs, being currently under discussion. The
29 proposed pathway referring to Article 43(3) and making use of the existing software-related codes³ to
30 include AI Act requirements enables notified bodies to assess MDAI in an efficient manner without
31 compromising due scrutiny. Although Article 30 foresees designation of AI NBs in a horizontal manner
32 (i.e. across different business areas), this pathway may be too time consuming and complicated to
33 establish enough NBs for MDAI to meet industry demands in a timely manner. However, we observe
34 with some concern that not all member states might be ready on 02 August 2025 with established
35 national implementing laws and designating authorities. Delays are likely to cause a shortage of
36 designated NBs, when the high-risk scope of the AI Act becomes applicable on 02 August 2027.

37 Conformity assessment of MDAI will follow the procedures established under MDR and IVDR (Art.
38 43(3)). This will allow both, manufacturers and NBs to focus on new requirements of the AI Act. As
39 there is already substantial expertise and experience on both sides, we foresee the challenges of the
40 conformity assessment of high-risk MDAI to be of administrative and legal nature.

41 Particularly, a common understanding of the AI Act's definitions of 'AI system' and 'safety component'
42 is crucial to the scope. Likewise, the terms 'substantial modification' and 'significant change' require
43 clear definitions for life-cycle management of MDAI. With the Commission guideline on the AI system
44 definition, we see an important first step. Yet, additional clarity will be needed for all stakeholders,
45 especially when it comes to specific knowledge- or logic based approaches as either AI or non-AI
46 models.

47 While safety and performance remain central for medical devices, the AI Act makes distinct provisions
48 for the compliance with fundamental rights like privacy, data protection, non-discrimination, and
49 human dignity. These principles need to be incorporated into manufacturers' risk assessments and
50 products. Further guidance on fundamental rights in this context will help harmonise approaches and
51 expectations. Also, embedding principles of fundamental rights into standards (compliance by design)
52 may support effective compliance with the AI Act's requirements.

53 The AI Act and MDR/IVDR have requirements for post-market surveillance and vigilance. This means
54 that there is some risk relating to duplication of administrative efforts as well as compartmentalisation
55 of information. By ensuring interoperability between surveillance systems and databases, full oversight
56 may be maintained and administrative burden reduced.

57 It is a widely acknowledged fact that the availability of high-quality data is pivotal for training,
58 validation and testing of AI systems. While this challenge is foremost one of developers of AI systems,
59 NBs will have responsibilities regarding conformity assessments through GDPR-compliant access to
60 manufacturers' data sets and, if necessary, also to high-quality, well-documented, and independent
61 datasets for additional testing of MDAI. Whilst a fully implemented European Health Data Space (EHDS)

³ Regulation 2017/2185



The European Association of
Medical devices Notified Bodies

Team-NB Position Paper

62 may be particularly helpful we must acknowledge that this won't be the case for the applicability of
63 the high-risk scope of the AI Act. Finding timely and practical solutions to these data challenges will be
64 necessary.

65 Timelines for the finalisation of harmonised standards under the AIA are already delayed. Efficient
66 development and conformity assessment of innovative MDAI relies greatly on their availability.
67 Although their absence may not be insurmountable given the framework which considers state of the
68 art, not being able to rely on harmonised standards will cause compliance issues and costs for
69 developers and assessors of MDAI such as notified bodies.

70 In summary, the implementation of the AI Act in the medical device sector hinges on successful
71 completion of these challenges. Clear definitions, timely national laws, availability of designating
72 authorities, robust data security, integrated reporting systems, consideration of fundamental rights,
73 and availability of harmonized standards are essential for smooth and effective conformity
74 assessments and ensuring access to advanced medical technology.

75



The European Association of
Medical devices Notified Bodies

Team-NB Position Paper

76 **Designation of Notified Bodies**

77 Background: The AI Act offers different routes for conformity assessment bodies to assess AI systems.
78 First, conformity assessment bodies can apply for full designation according to Article 30 which is
79 primarily to be understood in context of Article 6(2) (Annex III). There are several provisions in the AI
80 Act to make this pathway as smooth as possible (e.g. avoiding duplicate documentation of
81 requirements for notified bodies) Yet, it is to be expected that this process will take considerable time.
82 Designating authorities have to be established on member state level by 02 August 2025 (Article 70
83 (2)). Second, Article 43 (3) allows the conformity assessment of high-risk AI systems under the
84 respective specific Union harmonisation legislation (listed Annex I as referred to in Article 6(1), incl.
85 MDR and IVDR) covering specific requirements of the AI Act as part of the established conformity
86 assessment procedures. Both designation routes (1. Full designation according to AI Act (Annex III); 2.
87 Extended designation under other Union harmonisation legislation (Annex I)) require national
88 implementing laws in all member states. Formally, notified bodies may only pursue a designation for
89 assessing high-risk AI systems when these national laws are established as required by the AI Act.

90 Opinion: We greatly welcome the discussions on member state level to establish a pragmatic way to
91 implement the AI Act following the second option described above. Particularly, the proposed
92 approach to extend the scope of software codes established by Regulation 2017/2185 to cover
93 additional requirements of the AI Act appears to offer a fast and easy route to enable conformity
94 assessment of AI systems in medical devices without compromising the scrutiny on notified bodies.

95 With increasing concern, we observe recent discussions, that designations according to this second
96 option shall not be possible. This position is in contradiction to Article 43(3) and bears the imminent
97 risk of notified bodies not being able to pursue their obligations in a timely manner. A limited
98 availability of NBs at the time of applicability of the AI Act for medical devices is likely to be the
99 consequence.

100 Furthermore, we'd like to highlight, that not all member states are diligently proceeding with national
101 implementing acts, that are required to establish a legal basis for any conformity assessment according
102 to the AI Act. We acknowledge that potential delays are due to specific national political constellations.
103 Yet, this poses a substantial risk to innovative AI systems throughout the Union and a balanced
104 competition between notified bodies. Notified bodies are preparing for the applicability of the AI Act,
105 but it should be noted that actual designation procedures will require additional time after national
106 implementing laws have come into force. Consequently, we see the risk of notified bodies not being
107 readily available well in advance of the applicability of the high-risk scope of the AI Act (i.e. 02 August
108 2027). Depending on the practical interpretation of the overall scope of the AI Act (see 'Definitions'),
109 we'd like to make aware of the inherent risk of a major disruption of the medical device software
110 (MDSW) market.

111 We strongly believe that the swift designation of notified bodies is pivotal to the implementation of
112 the AI Act. Any delays on European or member state level will likely cause backlogs in assessing
113 innovative medical devices and thus prevent healthcare professionals and patients having access to
114 cutting edge medical technology.



The European Association of
Medical devices Notified Bodies

Team-NB Position Paper

115 **Conformity Assessment Procedure**

116 Background: In accordance with Article. 43 (3), an integrated process for conformity assessment of
117 medical devices including or consisting of AI systems is required.

118 Opinion: AI methodologies in medical technology are readily being used and assessed for years.
119 Consequently, notified bodies and industry have gathered considerable experience and built expertise
120 in this field. AI systems are considered medical device software and are assessed accordingly within
121 the current framework, i.e. for conformity with relevant GSPRs. Given the existing capacities for the
122 assessment of AI in medical devices and ongoing efforts to increase these to meet industry demand,
123 we expect a smooth building of capacities on the AI Act by notified bodies.

124 When the high-risk scope of the AI Act is applicable on 02 August 2027, medical devices using AI in
125 accordance with Art. 6(1) are to be compliant with the AI Act. Particularly the following points will
126 require additional attention by manufacturers as well as notified bodies (list is not exhaustive):

- 127 • Ensure a high level of protection of fundamental rights in addition to health and safety (Art. 1)
- 128 • Risk and quality management systems compliant with Articles 9 and 17 respectively,
- 129 • Stringent provisions ensuring human oversight over AI systems need to be in place (Art. 14),
- 130 • Appropriate logging needs to be in place (Art. 12)
- 131 • If deemed necessary, NBs may access training, validation and testing datasets of AI systems
132 Annex VII, 4.3),
- 133 • If technical documentation does not provide clear evidence for compliance with the AI Act,
134 NBs may carry out additional testing on their own or commission tests (Annex VII, 4.4 and 5.3),
- 135 • If tests and audit are not sufficient, NBs may access source code of AI systems (Art. 74 (13)(a)).

136 Although the overall requirements for MDAI under the MDR/IVDR regulatory framework are already
137 at a considerably high level and ensure a robust compliance framework, the AI Act adds substantial
138 requirements (see above). Yet, we believe, that administrative aspects represent a greater challenge
139 to the implementation of the AI Act. Thus, it will be crucial to follow a well-coordinated approach
140 between member states, that are in charge of notified body oversight, and the European Commission.

141

142 **Definitions**

143 Background:

144 Article 3

145 (1) 'AI system' means a machine-based system that is designed to operate with varying levels of
146 autonomy and that may exhibit adaptiveness after deployment, and that, for explicit or implicit
147 objectives, infers, from the input it receives, how to generate outputs such as predictions, content,
148 recommendations, or decisions that can influence physical or virtual environments;'

149 [...]



The European Association of
Medical devices Notified Bodies

Team-NB Position Paper

150 (14) ‘safety component’ means a component of a product or of an AI system which fulfils a safety
151 function for that product or AI system, or the failure or malfunctioning of which endangers the health
152 and safety of persons or property;

153 [...]

154 (23) ‘substantial modification’ means a change to an AI system after its placing on the market or putting
155 into service which is not foreseen or planned in the initial conformity assessment carried out by the
156 provider and as a result of which the compliance of the AI system with the requirements set out in
157 Chapter III, Section 2 is affected or results in a modification to the intended purpose for which the AI
158 system has been assessed;

159 [...]

160 Opinion: These definitions are crucial for understanding the scope of the AI Act, as they define (1) the
161 overall scope of the AI Act, (14) the scope of high-risk AI systems, and (23) the need for post-market
162 change reporting.

163 (1) We appreciate the adoption of a wording for the ‘AI system’s’ definition that follows the OECD
164 definition. We believe this will allow for an easier global harmonisation and help avoiding
165 diverging positions on the nature of AI between major markets. For good reasons, this covers
166 a wide scope to allow for interpretation according to future technological developments in the
167 field of AI.

168 We believe the clarifying guidance⁴ by the European Commission is of great value for a
169 consistent understanding of the legal definition of AI systems. Also, we appreciate the need
170 for flexibility for case-by-case decisions which is foreseen by this guideline. However, the pivot
171 of the AI Act’s definition of AI systems ‘inference’ might require additional elaboration, e.g. by
172 more detailed and sector specific guidance making use of examples. Unclear constraints of the
173 definition might well cause an overflow of the definition to include entities that are
174 traditionally not considered AI systems (e.g. many logic- or knowledge-based systems).

175 (14) Another critical aspect for understanding the impact of the AI Act is the term ‘safety
176 component’. In the current discourse, we observe a certain divergence of opinions on what
177 uses of AI systems qualify as safety component. Our understanding of the term is that the AI
178 system as a part of a medical device always bears the risk of ‘endanger[ing] the health and
179 safety of persons’. Obviously, this will always be the case, if the AI system *is* the medical device.
180 The definition of a ‘safety component’ in the AI Act, allows the conclusion that a potential
181 malfunction of an AI system always constitutes a ‘device deficiency’ according to MDR, Art. 2
182 (59). However, a more pragmatic interpretation appears to be more appropriate for practical
183 reasons.

184 Since the high-risk scope of the AI Act and also medical ‘device deficiencies’ directly relate to
185 this term, clarifying guidance would be highly appreciated.

⁴ Commission Guidelines on the definition of an artificial intelligence system established
by Regulation (EU) 2024/1689 (AI Act), C(2025) 924 final, published 06 February 2025



The European Association of
Medical devices Notified Bodies

Team-NB Position Paper

186 (23) Likewise, understanding of ‘substantial modifications’ will be needed; according to recital 177,
187 it can be understood as equivalent to the term ‘significant change’ which is used by AI Act and
188 MDR/IVDR. However, ‘significant changes’ lack a formal legal definition in the latter context
189 and have been subject to disputes. The guidance required by Article 96 (AIA) could be a good
190 opportunity to establish a harmonised interpretation of ‘substantial modifications’ and
191 ‘significant changes’.

192 We’d like to highlight, that the ease of implementation of the AI Act greatly depends on the
193 interpretation of the above mentioned and other definitions. Notably, this will affect the number of
194 products requiring an extended conformity assessment covering the requirements of the AI Act or re-
195 assessment following changes of the device.

196

197 **Fundamental Rights**

198 Background: In Article 1, the AI Act makes a very prominent reference to the protection of fundamental
199 rights, with reference to the Charter⁵ including democracy, the rule of law and environmental
200 protection. While ensuring a high level of safety and health is already the focus of MDR and IVDR the
201 topic of protection of fundamental rights is an additional focus provided by the Act (EU) 2024/1689.

202 Opinion: Though for medical devices the aspects of safety and health appear generally to be most
203 important, fundamental rights (e.g. privacy, data protection, non-discrimination, or human dignity and
204 autonomy) will also have to be considered. Medical device manufacturers will need to adapt their
205 approaches and take in consideration the relevant factors in their risk assessments and in consequence
206 in their products. The guidance to be provided on the matter of fundamental rights will be an important
207 orientation for harmonising approaches and expectations.

208

209 **Reporting and Vigilance**

210 Background: The AI Act requires registration and safety reporting (e.g. registration acc. to Art. 71 AI
211 Act; safety reporting acc. to Art. 73 AI) in a similar fashion to MDR and IVDR. New databases and IT
212 systems are to be established under the oversight of the AI Office.

213 Opinion: Given the existing/to be established reporting routes under MDR/IVDR, the new
214 requirements of the AI Act bear the risk of duplicating administrative efforts and partitioning
215 information in independent reporting systems. In line with the approach to be used for conformity
216 assessments (see above), it would be highly desirable that reporting requirements of the AI Act are
217 fully integrated into the existing routes for medical devices. If an integration is not feasible and to avoid
218 partitioned vigilance information spread over several systems, interoperability of respective databases
219 (i.e. information mirrored to the respective other system) is a minimum requirement. In this way, it
220 would be guaranteed both regulatory frameworks having full oversight and not missing any critical
221 information.

⁵ CHARTER OF FUNDAMENTAL RIGHTS OF THE EUROPEAN UNION ([2016/C 202/02](#))



The European Association of
Medical devices Notified Bodies

Team-NB Position Paper

222 **Data and Data Governance**

223 Background: Availability of high-quality data is key for the quality of an AI model or system and is
224 therefore important in the development of many AI devices, most relevantly those employing machine
225 learning techniques. Consequently, the EU AI Act imposes various requirements on data and data
226 governance that directly impact conformity assessments of high-risk AI systems (Art. 10). E.g. ‘Training,
227 validation and testing data sets shall be relevant, sufficiently representative, and to the best extent
228 possible, free of errors and complete in view of the intended purpose. [...]’ (Art. 10(3)). To verify the
229 adequacy of the data used by developers, conformity assessment may require access to those in many
230 cases. Additionally, NBs must verify that AI system providers have implemented robust data
231 governance practices, including traceability and compliance with GDPR (Regulation (EU) 2016/679).

232 For the purpose of conformity assessment, notified bodies may require full access to training,
233 validation and testing data sets used by the manufacturer (Annex VII, 4.3). Also, additional testing of
234 AI systems by notified bodies may be required during assessment of technical documentation (Annex
235 VII, 4.4). The availability of high-quality, well-documented and independent data sets is essential for
236 such additional tests.

237 Opinion: We see two critical challenges with regard to data for Notified Bodies:

- 238 • In many cases, NBs will have to access training and testing datasets used by developers to
239 assess conformity of devices. This will require appropriate data security measures and
240 compliance with GDPR by both, medical device manufacturers and notified bodies. E.g.
241 security of data when accessed remotely by or being transferred to NBs must be ensured as
242 well as appropriate data protection agreements have to be in place that allow this access by
243 a third party. Practically, this might prove to be demanding for both sides.
- 244 • The availability of a sufficient amount of high-quality, well-documented and independent
245 datasets for notified bodies will be a challenge. For example, the scarcity of this resource
246 might cause problems for the independence of data from training data. To enable notified
247 bodies to fulfil their responsibilities, easy access to data will be essential for verifying safety,
248 performance, and fairness of AI-driven medical devices. We’d like to note that currently no
249 easy pathway is established to guarantee access to independent test datasets.
250 We expect the implementation of the European Health Data Space (EHDS)⁶ will offer at least
251 a partial solution for the increased demand for health data. However, the EHDS framework
252 will not be readily available for the applicability of the high-risk scope of the AI Act, so that
253 availability of data might pose a considerable barrier to the certification of AI in or as medical
254 devices. Thus, finding pragmatic solutions to this problem should be a priority for all
255 stakeholders.

⁶ Regulation (EU) 2025/327 of the European Parliament and of the Council of 11 February 2025 on the European Health Data Space and amending Directive 2011/24/EU and Regulation (EU) 2024/2847, <https://eur-lex.europa.eu/eli/reg/2025/327/oj/eng>



The European Association of
Medical devices Notified Bodies

Team-NB Position Paper

256 Despite these challenges, it shall be noted that NBs' focus will remain on the assessment of quality
257 management systems and technical documentation of MDAI (AI Act, Annex VII); and that further
258 testing or scrutiny of data sets is not required by default.

259

260 **Standardisation**

261 Background: According to Article 40 of the AI Act, harmonised standards for AI systems are to be
262 drafted. The European Commission issued a first request for such activities already in May 2023 to CEN
263 and CENELEC.⁷

264 Opinion: Although standardisation requests have been made in advance, a delayed availability of
265 harmonised standards is to be expected. Although such are not strictly necessary for the
266 implementation, they would allow for a presumption of conformity with the provisions of the AI Act
267 and hence an easier development and conformity assessment of devices including an AI system. If
268 harmonised standards are not available well in advance of the applicability of the high-risk scope of
269 the AI Act, conformity assessment will have to rely on a comparison with the state-of-the-art which
270 will be more resource consuming than demonstrating compliance with a harmonised standard.

271 It should be noted that currently available horizontal standards on AI systems are not harmonised with
272 the AI Act and are neither adequate for presumption of conformity nor may be considered state-of-
273 the-art in their entirety.

274

275 **Version history**

276 This document is a full revision (Version 2) of the Team-NB Position Paper "European Artificial
277 Intelligence Regulation" (06 Oct 2021) and supersedes the position "The designation of notified bodies
278 under the upcoming Artificial Intelligence Act" (16 Dec 2022).

⁷ [C\(2023\)3215](#) – Standardisation request M/593, COMMISSION IMPLEMENTING DECISION of 22.5.2023 on a standardisation request to the European Committee for Standardisation and the European Committee for Electrotechnical Standardisation in support of Union policy on artificial intelligence